

# EC DECLARATION OF CONFORMITY

without the participation of a Notified body - diagnostic medical devices *in vitro*

Pursuant to Section 13(2) of Act No 22/1997 Coll., on the Technical Requirements for Products, and on a change and addition to certain laws, as amended, and pursuant to Directive 98/79/EC of the European Parliament and of the Council of 27 October 1998 on *in vitro* diagnostic medical devices, the requirements of which have been adopted by Government Regulation 56/2015 Coll., on Technical Requirements for *in vitro* Diagnostic Medical Devices, as amended,

MANUFACTURER

**GeneProof a.s., Vídeňská 101/119, Dolní Heršpice, 619 00 Brno, Czech Republic**  
**www.geneproof.com**

hereby declares that following product

## GeneProof *Bordetella pertussis/parapertussis* PCR Kit

**Medical device type:** *in vitro* diagnostic medical device  
**Classification:** other diagnostic medical devices  
**Intended purpose of the device:** The kit is intended for qualitative diagnostics and aid to diagnosis of *Bordetella pertussis* and *Bordetella parapertussis* from aspirate, sputum and swab by Polymerase Chain Reaction (PCR) method. The intended users are professional personnel of clinical laboratories.  
**Variants:** BP/ISEX/025, BP/ISEX/100

complies with the basic requirements of Annex No. 1 of the Directive 98/79/EC of the European Parliament and of the Council and the Czech Government Regulation No. 56/2015 Coll. and under normal use it is safe and effective for its intended purpose. The manufacturer has taken measures assuring compliance of all medical devices introduced into the market with their technical documentation and with the essential requirements.

Following currently valid version of the standards were applied to meet essential requirements:

ČSN EN ISO 13485 ed.2:2016	Medical device – QMS – Requirements for regulatory purposes
ČSN EN ISO 14971:2020	Medical device – QMS – Application of risk management to medical devices
ČSN EN ISO 18113-1:2012	<i>In vitro</i> diagnostic medical devices -- Information supplied by the manufacturer (labelling) -- Part 1 Terms, definitions and general requirements
ČSN EN ISO 18113-2:2012	<i>In vitro</i> diagnostic medical devices -- Information supplied by the manufacturer (labelling) -- Part 2 <i>In vitro</i> diagnostic reagents for professional use
ČSN EN ISO 15223-1:2017	Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied, Part1: General requirements
ČSN EN ISO 23640:2016	<i>In vitro</i> diagnostic medical devices – Evaluation of stability of <i>in vitro</i> diagnostic reagents

Procedure described in Annex No. 3 of the Directive 98/79/EC of the European Parliament and of the Council and the Czech Government Regulation No. 56/2015 Coll. were used to evaluate the basic characteristics of the product by the designated method.

Brno, May 18, 2022

**Kamil ŠPLÍCHAL**  
**Quality Assurance/Regulatory Affairs**  
**Chief Quality and Regulatory Affairs Officer**  
(Name, position and signature of authorized person)



Manufacturer's stamp:

