




# FlashDx<sup>®</sup> Canine RP 1.0 (Respiratory Panel)

<b>Test Reagent Kit</b>	Canine Respiratory Panel 1.0	
<b>Catalog Number</b>	CAN-P5-10	
<b>Instrument</b>	FDx-100	
<b>Technology</b>	qPCR-Microarray	
<b>Targeted Pathogens</b>	<b>CDV</b> Canine Distemper Virus	<b>CAV-2</b> Canine Adenovirus Type-2
	<b>CPIV</b> Canine Parainfluenza Virus	<b>Bb</b> <i>Bordetella bronchiseptica</i>
	<b>MC</b> Mycoplasma cynos	
<b>Batch or On-Demand</b>	On-Demand	
<b>Minimum Batch Size</b>	1	
<b>Sample Types</b>	<b>Specimen:</b> Conjunctival Swab, Oropharyngeal Swab <b>Transport Media:</b> UTM/VTM* or saline (0.9%)	
<b>Sample Extraction</b>	Fully automated/integrated	
<b>Precision Pipetting</b>	Not required	
<b>Turnaround Time</b>	~45 minutes, positive results as fast as 30 minutes with "Early Call Assays" (ECA)	
<b>Hands-on Time</b>	< 1 minute	
<b>Control</b>	Internal Process Control	
<b>Kit Storage</b>	2-8 °C	
<b>Commercial Controls</b>	See FlashDx Canine RP1.0 Package Insert or contact FlashDx Customer Support	

\* Sample transport media with guanidine buffer is not compatible.



## USA

**Phone** +1 (650) 260-3737  
**Address** 303 Twin Dolphin Drive  
Suite 600  
Redwood City, CA 94065

## CHINA

**Phone** +86 755 8696 5752  
**Address** B6-4C, China Merchants  
Guangming Science Park  
Guangming District, Shenzhen

[www.flash-dx.com](http://www.flash-dx.com)  
[info@flash-dx.com](mailto:info@flash-dx.com)

Ver. 0.1