




FlashDx[®] Feline RP 1.0 (Respiratory Panel)

Test Reagent Kit	Feline Respiratory Panel 1.0						
Catalog Number	FEL-P6-10						
Instrument	FDx-100						
Technology	Microarray-qPCR						
Targeted Pathogens	<table><tr><td>FHV-1 Feline Herpes Virus-1</td><td>FCV Feline calicivirus</td></tr><tr><td>MF <i>Mycoplasma felis</i></td><td>Bb <i>Bordetella bronchiseptica</i></td></tr><tr><td>Flu-A Feline influenza A</td><td>CF <i>Chlamydia felis</i></td></tr></table>	FHV-1 Feline Herpes Virus-1	FCV Feline calicivirus	MF <i>Mycoplasma felis</i>	Bb <i>Bordetella bronchiseptica</i>	Flu-A Feline influenza A	CF <i>Chlamydia felis</i>
FHV-1 Feline Herpes Virus-1	FCV Feline calicivirus						
MF <i>Mycoplasma felis</i>	Bb <i>Bordetella bronchiseptica</i>						
Flu-A Feline influenza A	CF <i>Chlamydia felis</i>						
Batch or On-Demand	On-Demand						
Minimum Batch Size	1						
Sample Types	Specimen: Conjunctival Swab, Oropharyngeal Swab Transport Media: UTM/VTM* or saline (0.9%)						
Sample Extraction	Fully automated/integrated						
Precision Pipetting	Not required						
Turnaround Time	~45 minutes, positive results as fast as 30 minutes with "Early Call Assays" (ECA)						
Hands-on Time	< 1 minute						
Control	Internal Process Control						
Kit Storage	2-8 °C						
Commercial Controls	See FlashDx Feline RP1.0 Package Insert or contact FlashDx Customer Support						

* Sample transport media with guanidine buffer is not compatible.



USA

Phone +1 (650) 260-3737
Address 303 Twin Dolphin Drive
Suite 600
Redwood City, CA 94065

CHINA

Phone +86 755 8696 5752
Address B6-4C, China Merchants
Guangming Science Park
Guangming District, Shenzhen

www.flash-dx.com
info@flash-dx.com

Ver. 0.2